



SECTION A-A

The Recirculating Lab Faucet comes with the Recirculating Module and Tube temporarily assembled.

STEP 1. Remove the Recirculating Module and Recirculating Tube from the Through Wall Nipple.

STEP 2. Mount the faucet in the desired location through a 1" diameter hole. Tighten to wall with the Ring Nut supplied.

STEP 3. Insert the Recirculating Tube into and through the Through Wall Nipple until the 3/8" male NPT thread mates with the 3/8" female NPT thread in the inlet connection of the Through Wall Nipple. Tighten by hand plus 1/4 wrench turn.

STEP 4. Connect the supply line to the 3/8" female NPT inlet connection.

STEP 5. Connect the recirculating return line to the 3/8" female NPT outlet connection.

STEP 6. Pressure test all connections.

Note: A high density PTFE tape is the recommended thread sealant.

2	Polypropylene, Natural
3	PVDF, Natural

Designer		
Review		
Engineer		

TITLE:	
Recirculating Lab Faucet Wall Mount, Duraline Control Valve	



DWG. NO.	REV
LG-WDR-(R, L or F)(2 or 3)	

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES 2950 Airway Avenue, Bldg. C5...Costa Mesa, CA 92626

SCALE: N/A	Format: SD	SHEET 1 OF 1
------------	------------	--------------